

Errors in Laboratory Medicine

Mario Plebani, M.D.

Chief, Department of Laboratory Medicine
University-Hospital of Padova, Italy



To erris human:

huilding a safer health system



Kohn L.T., Corrigan J.M., Donaldson M.S., Eds. 1999




What About Use




- Risk of death from avoidable injury
 - -2,917 per 1,000,000













- The number of deaths that are reportedly due to medical errors is disturbingly high.
- The IOM report suggests that more Americans are killed in US hospital every 6 months than died in the entire Vietnam War.
- If true, the healthcare system is a public health menace of epidemic proportions.



◆ There is a shortage of scientifice evidence for documenting the types of laboratory errors and their frequency and few studies consider the clinical impact of laboratory errors on medical and economic outcomes.

Plebani M, Bonini P. BMJ 2002; 324: 423-




◆ The literature on errors in laboratory medicine is scarce, both for the insufficient attention paid to the problem, and for practical difficulties in reporting and measuring the number of errors.

Bonini P, Plebani M, Ceriotti F, Rubboli F



- Study design.
- Errors in laboratory or errors in laboratory medicine?
- Difficulties in identifying all types of errors.
- Lack or changing of Gold standards.



- Most of the studies focus simply on analytical errors, which represent only a percentage of the errors in the total testing process (TTP), which includes all pre-, intra-, and post-analytical phases.
- Other studies are based on methodologies, such as the split-specimen design, that are insensitive to some steps of the TTP.



Laboratory errors in the Total Testing Process

Total errors patients involved: 363

• Pre-analytical steps 218 (45.5%)


• Analytical steps 35 (7.3%)

• Post-analytical steps 266 (47.2%)


Ross JW, Boone DJ, 1991



◆ Apart from a reluctance in reporting their own errors, it is extremely difficult for laboratories to identify all errors because many errors will neither produce detectable abnormal results nor raise questions for the user.



- The reported errors are only the tip of the iceberg because, as stated by Goldschmidt,
- a) 75% of errors produce results within the reference intervals.
- b) 12.5% produce wrong results that are so absurd that they are not considered clinically
- C) the remaining 12.5% of errors may have effect on patient health.





The new pathophysiologic insights and the development of highly specific and sensitive laboratory tests have changed the relationship between laboratory information and the gold standards.



One identified error every

33-50 events

50-100 events

330 events

1000 events

2237 events

8300 laboratory results or

2000 patients

900 patients

214 laboratory results

164 laboratory reports

283 laboratory results

McSwiney and Woodrow ('69)

Souverijn et al. ('80)

Chambers et al. ('86)

Boone ('90)

Witte et al. ('97)

Lapworth and Teal ('94)

Nutting et al. ('96)

Plebani and Carraro ('97)

Stahl et al. ('98)

Hofgartner and Tait ('99)



Review of the literature on laboratory errors

SACIES INTERCEDICAL					
ector of the laboratory	Lapworth and Teal (10) Clinical chemistry	Goldschmidt and Lent (7) Whole laboratory	Nutting et al. (36) Primary care	Plebani and Carraro (8) Stat laboratory	Stahl et al. (37) Whole laborator
lata collection period	1 year	6 years	6 months	3 months	3 years
lo. of test	~997 000	Nda	ND	40 490	676 564b
lo. of patients	~249 000	ND	160 714	ND	ND
lo. of errors	120	133	180	189	4135 ^b
requency	0.05% of patients		0.11% of patients	0.47% of test results	0.61% ^b of test
Preanalytical phase Analytical phase	31.6% 31.6%	53% 23%	55.6% 13.3% overall (4.4% if referral laboratory; 40% if POCT)	68.2 13.3%	75% 16% ^b
Postanalytical phase Multiple phases	30.8% 6%	24%	30%	18.5%	9% ^b
dentification errors	41 (34%)	77 (58%)	ND	5 (2.6%)	ND



♦ It is possible, even probable, that the most frequent pre-analytical errors are represented by an inappropriate choice of laboratory tests or panel of tests, and that most post-analytica errors derive from inappropriate interpretation and utilization of laboratory results.



ACAP Q-Probes Study of Requisition Order Entry Accuracy

	4 V	70
• Total test requisitions examined	114,934	
 Total order entry errors 	5,514	4.8
• Test(s) ordered but not performed	1,658	1.4
• Test(s) performed but not ordered	1,221	1.1
• Physician name discrepancies	2,130	1.9
 Test priority errors 	934	0.8

Arch Pathol Lab Med 1999;123:1145-50



Types of preanalytical errors registered during the year 200 at the Laboratory of San Raffaele Hospital

	No. Of missing results		
Type of error	Inpatients	Outpatients	
Hemolyzed sample	8494	256	
Insufficient sample	3256	102	
Incorrect sample	1824	289	
Clotted sample	792	80	
Incorrect identification	287	2	
Lack of signature (blood group)	266		
Empty tube	238	8	
Lack or wrong compilation of the			
accompanying module	120	-	
Sample not on ice	75	6	
Tube broken in the centrifuge	57	36	
Test not reserved	31	-	
Urine not acidified	24	-	
Open container	20	13	
Module without signature	14	-	
Urine volume not indicate	5	-	
Total	15 503	792	



Laboratory errors in the CAP Q-Proby of blood bank practices

Total errors

64,000

Pre-analytical steps

52%

Analytical steps

5%

• Post-analytical steps

43%

Bachner P, Boone DJ, et al., 1991



Has the unacceptable result rate improved over time?

ppm

- **♦ Belk and Sunderman (1947)**
- ◆ College of American Pathologists (1996)
- Plebani and Carraro (1997)
- ♦ Witte D.L. et al. (1997)
 - ("pure" analytical errors)

162,116


12,904

4,700

447



Mistakes and patient outcomes



J.W. Ross and

P.A. Nutting et al.

M. Plebani, P. Carraro

D.J.Boone

<i>Number</i>	Effecton patient care (%)		
336	30		
180	27		
189	26		

Risk of inappropriate care (%)
7
12
6.4



Overall rates of unsatisfactory and unsuccessful performance

Facility type	Total analyte challenges	Unsatisfactory % (n°)	Unsussessf % (n°)
Physiciansoffice laboratories	656	21.5 (141)	4.4 (29)
Physicians office laboratories using medical technologists	662	14.0 (93)	1.8 (12)
Non-physicians'office laboratories	2991	8.1 (243)	0.9 (26)

Hurst J., JAMA 1998;279:468-71.



Odds ratio of unsatisfactory proficiency testing event performance

(all other testing sites vs hospital and independent laboratories)

Analyte

Odds ratio confidence (95% interval)

All analytes

2.89

Potassium

7.51 (6.59-8.54)

•Theophylline

5.5 (4.64-6.64)

•Hemoglobin

4.56 (4.08-5.09)

•Uric acid

4. 32 (3.78-4.94)

Total bilirubin

4.28 (3.78-4.85)

Stull TM et al., JAMA 1998;279:463-7.



In order to identify the most critical steps in the TTP and to set up a plan for a corrective strategy we have to make a distinction between:

- a) Errors exclusively inside the laboratory.
- b) Laboratory errors caused by organizational problems outside the laboratory.
- c) Errors at the laboratory-clinical interface.





ERRORS IN LABORATORY MEDICINE TENTATIVE CLASSIFICATION

Errors exclusively inside the laboratory:

- Analytical errors
- Acceptance of improper specimens
- Sample mismatch during the analysis
- Undue increase in TAT
- Mistakes or failures in reporting




Laboratory errors caused by organizational problems outside the laboratory should be defined

"patient investigation errors"

 (e.g. sample-patient mismatch during the blood withdrawal performed by nonlaboratory personnel).

TYPES OF WRISTBAND ERRORS





WRISTBAND ERROR RATES (%) - CAP Q-TRACKS STUDIES-

1991

5.5

(10% of participants with error rates ≥ 10.9%)

1993

8.4

(hospitals with fewer tha 200 beds)

1995

7.4


♦ 1999 first quarter

7.4

◆ 1999 eight quarter 3.05



The Laboratory Can Reduce Patient Identification Errors



- Phlebotomy staff can monitor
- Interventions can reduce discrepancies
- Reduced discrepancies reduces disasters

Arch Pathol Lab Me 1993;117:573-57



DENTIFICATION OF THE PATIENT



BOND JAMES dd/mm/yy M RAF007 123456789012 990088776655 card antib aspir

Applicazione protetta da brevetta




DRIVE - Drug in Virtual Enterprise

DRIVEAPPROACH

-for a safe, smart Hospital -







ERRORS IN LABORATORY MEDICINE: TENTATIVE CLASSIFICATION

Errors at the laboratory-clinical interface:

- Appropriateness in test request.
- ◆ Appropriateness in test interpretation.
- ◆ Appropriateness in test utilization.



PHYSICIAN SURVEY ON CBC/DIFF REPORTS: STUDY RATIONALE

- CBC/Diff reports have become increasingle long and complex.
- Unnecessary information provided to clinicians may impede their comprehension of essential results.
- Simplification of reports might improve comprehension of the results and therebased reduce the potential for medical errors.

Linda M Sandhau, AACC 200






The CBC has become a monster!



Errors Occur When All Defenses Fail









HAZARDS

MISTAKES AND SLIPS:

- Errors of individuals
- Violations
- Diagnostic mishap

ORGANIZATIONAL FAILURES:

- Procedures/Processes
- Cultural constraints
- Legal and regulatory rules
- Failures in communications

Injuries and adverse incidents



Errors in Total Testing Process

(Procedures, Processes)

Adverse incidents

Errors NOT associated with negative patient outcomes

Errors associated with negative patient outcomes

Injuries associated with errors

Events included in error-focused approach


Events included in adverse Incidents-focused approach



ERRORS IN LABORATORY MEDICINE what should we do?

- There is a need for better definition of laboratory errors and their causes
- ♦ In fact, we can agree that laboratory error may be defined as "any defect from ordering tests to reporting results and appropriately interpreting and reacting on these"

but



It is important to classify laboratory errors by relating them to their real or potential effects on patient outcomes, allowing definition of the relevance of the error itself.

(e.g. a hemolyzed sample is probably less problematic than sample mismatching or a TAT too long in a critical situation).




A standard for laboratory error detection and reporting needs to be defined, and an accurate analysis of the risk of errors in the clinical laboratory needs to be performed.




ERRORS IN LABORATORY MEDICINE: what should we do?

It is important to define ways to decrease laboratory errors and to possibly avoid completely those with a real or potentially significant negative effect on a patient's health.



An appropriate error detection program and adequate measures for error reduction that quantify the effects of these measures and evaluate whether the reduction can be considered satisfactory are critical.



Another fundamental step is to create a culture in which the existence of risk is acknowledged and injury prevention is recognized as everyone's responsibility.









Scientific Societies
Laboratory Professionals
Clinicians
Patients
Public

ERRORS REDUCTION MUST BEGIN AND ENL WITH RELATIONSHIPS

McNutt RA et al., JAMA 2002;287:1997-2001